

DiAgam

CATALOGUE 2019

COBAS Line





Company profile

MISSION STATEMENT

«To have a perfect control of Development, Manufacturing process & Selling of In Vitro Diagnostic liquid reagents with a strong focus on our Customer vision»

Your R&D Partner

DiAgam manufactures **reagents for immunoturbidimetry**.

Our company specializes in high technology R&D, based on immunology technologies. DiAgam is the creator of liquid laboratory reagents containing **gold particles**.

We are currently taking part in a project to develop **new nanotechnologies** based on the use of **colloidal particles** in the field of medical diagnosis. This activity has also succeeded in completing our products range with Cystatin C, Ferritin & Rheumatoid factor.

Manufacturer of ready-to-use reagents

All our products are developed and manufactured within the company.

Our reagents, controls and calibrators are liquid, stable and ready to use on **all opened biochemistry ana-lyzers even with barcodes** (Abbott - Alinity – Architect , Beckman Coulter – AU series, IDS-iSYS Multi-Discipline Automated System, Menarini – Zenit RA, Mindray BS series, Ortho - VITROS® 5,1 FS Chemistry, VITROS® 4600 Chemistry, VITROS® 5600 Integrated Systems, Roche - Modular®- COBAS C, Siemens - Advia® 1800 - Advia® 2400 - Atellica®, Thermo-Konelab®, etc...)

DiAgam is a member of the National Association of **European Diagnostic Manufacturers Association (EDMA)**. All reagents are labelled and DiAgam is certified **ISO 9001:2015** and **ISO 13485:2016**.

All calibrators and controls are traceable to the **International Standards** (WHO, IFCC, NIBSC. etc.)

Worldwide reliability

DiAgam is present all around the world. We sell directly in Belgium, France, Spain & Portugal. We also export to French Speaking African countries, Argentina, Brazil, Czech Republic, Greece, Hungary, India, Mexico, Morocco, Poland, Sri Lanka, Ukraine, Vietnam, etc.

Thanks to our **manufac-turing flexibility**, we can propose all our parameters in three different packagings: **Kits** especially dedicated to particular analyzers, **Vials** filled and **Bulk**, available under **DiAgam** label or in **OEM**.

We respect our philosophy by offering our cus-tomer, a **high quality of our products** with a most **competitive price** along with a continuous quest to **satisfy their needs**.



CATALOGUE 2019

ALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ALTUR-B00/COB	1 x 25,3 ml R1 + 1 x 18,2 ml R1 + 1 x 12,2 ml R2		230	36 months	serum or plasma	0.546 to 132.48 (g/l)
	ALTUR-H00/COB	2 x 25,3 ml R1 + 2 x 18,2 ml R1 + 2 x 12,2 ml R2		460			
	ALTUR-L00/COB	12 x 25,3 ml R1 + 12 x 18,2 ml R1 + 12 x 12,2 ml R2		2760			
	ALTUR-C00/COB	12 x 25,3 ml R1 + 12 x 18,2 ml R1 + 12 x 12,2 ml R2		2760			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 GLYCOPROTEIN ACIDE/OSOMUCOÏD

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	GPTUR-B00/COB	1 x 22,2 ml R1 + 1 x 14,3 ml R1 + 1 x 8,3 ml R		150	36 months	serum	0.143 to 2.42 (g/l)
	GPTUR-H00/COB	2 x 22,2 ml R1 + 2 x 14,3 ml R1 + 2 x 8,3 ml R		300			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 ANTITRYPSIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ATTUR-B00/COB	1 x 23,5ml R1 + 1 x 6,2 ml R2		100	24 months	serum or plasma	0.007 to 3.5 (g/l)
	ATTUR-H00/COB	2 x 23,5ml R1 + 2 x 6,2 ml R2		200			

CALIBRATORS

H	MPREH-001(-005)	1x1 or (1x5) ml	24 months
KIT 5 LEV	MPREK-000	5x1 ml	24 months

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 MICROGLOBULIN (URINE)

Contact us

CATALOGUE 2019

ALPHA-2 MACROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	MGTUR-B00/COB	1 x 23,5ml R1 + 1 x 6,4 ml R2	100	36 months	serum or plasma	0.026 to 4.96 (g/l)
	MGTUR-H00/COB	2 x 23,5 ml R1 + 2 x 6,2 ml R2	200			
	MGTUR-L00/COB	12 x 25,1 ml R1 + 12 x 17,2 ml R1 + 12 x 9,4 ml R2	2220			
	MGTUR-C00/COB	12 x 23,5 ml R1 + 12 x 6,2 ml R2	1200			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months
URINARY LOW	MGCU1-002	1x2 ml	12 months
URINARY HIGH	MGCU2-002	1x2 ml	12 months

ANTI STREPTOLYSIN O (ASO)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	ASLTX-B00/COB	1 x 23,5 ml R1 + 1 x 16,9 ml R1 + 1 x 6,8 ml R2	180	20 months	serum or plasma	19.99 to 864.85 (IU/ml)
	ASLTX-H00/COB	2 x 23,5 ml R1 + 2 x 16,9 ml R1 + 2 x 6,8 ml R2	360			

CALIBRATORS

H	ASREH-001(-005)	1x1 or (1x5) ml	12 months
---	-----------------	-----------------	-----------

CONTROLS

M	ASCON-002	1x2 ml	12 months
---	-----------	--------	-----------

ANTITHROMBIN III

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2			
KIT	A3TUR-B00/COB	1 x 22,2 ml R1 + 1 x 14,3 ml R1 + 1 x 8,3 ml R2	150	36 months	plasma	28.15 to 806.30 (mg/l)
	A3TUR-H00/COB	2 x 22,2 ml R1 + 2 x 14,3 ml R1 + 2 x 8,3 ml R2	300			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE 2019

APOLIPOPROTEIN A1

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	AATUR-B00/COB	1 x 22,2 ml R1 + 1 x 14,3 ml R1 + 1 x 10,4 ml R2	150	36 months	serum or plasma	0.029 to 400 (mg/dl)
	AATUR-H00/COB	2 x 22,2 ml R1 + 2 x 14,3 ml R1 + 2 x 10,4 ml R2	300			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x0.5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

APOLIPOPROTEIN B

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	ABTUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2	186	36 months	serum or plasma	0.019 to 4.0 (g/l)
	ABTUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2	558			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

BETA 2 MICROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	B2DGX-B00/COB	1 x 24,8 ml R1 + 1 x 16,4 ml R1 + 1 x 17,7 ml R2	240	15 months	serum or plasma	0,132 to 95 (mg/l)
	B2DGX-H00/COB	2 x 24,8 ml R1 + 2 x 16,4 ml R1 + 2 x 17,7 ml R2	480			
	B2DGX-L00/COB	12 x 24,8 ml R1 + 12 x 16,4 ml R1 + 12 x 17,7 ml R2	2880			

CALIBRATORS

KIT 5 LEV	B2REK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	B2COX-001	1 x 1 ml	18 months
M	B2CON-003	1x3 ml	18 months
L	B2COS-003	1x3 ml	18 months

BETA 2 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	BUDGX-B00/COB	Contact us		15 months	Urine	0.055 to 9.3 (mg/l)
	BUDGX-H00/COB					
	BUDGX-L00/COB					

CALIBRATORS

KIT 5 LEV	BUREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	B2CON-003	1x3 ml	18 months
L	B2COS-003	1x3 ml	18 months

CERULOPLASMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	CETUR-B00/COB CETUR-H00/COB	R1 1 x 23,5 ml R1 + 1 x 6,2 ml R2	120 240	36 months	serum	9.48 to 1051.20 (mg/l)
		R2 2 x 23,5 ml R1 + 2 x 6,2 ml R2				

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C3

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	C3TUR-B00/COB C3TUR-H00/COB	R1 1 x 20,2 ml R1 + 1 x 13,0 ml R1 + 1 x 7,7 ml R2	133 400	36 months	serum or plasma	0.015 to 4.31 (g/l)
		R2 3 x 20,2 ml R1 + 3 x 13,0 ml R1 + 3 x 7,7 ml R2				

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C4

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT	C4TUR-B00/COB C4TUR-H00/COB	R1 1 x 23,5 ml R1 + 1 x 15,4 ml R1 + 1 x 8,7 ml R2	170 510	36 months	serum or plasma	0.002 to 0.825 (g/l)
		R2 3 x 23,5 ml R1 + 3 x 15,4 ml R1 + 3 x 8,7 ml R2				

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE 2019

CRP GOLD LATEX

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CPDGX-B00/COB	1 x 20,2 ml R1 + 1 x 18,3 ml R2	150	16 months	serum or plasma	0.257 to 370.56 (mg/l)
	CPDGX-H00/COB	3 x 20,2 ml R1 + 3 x 18,3 ml R2	450			
	CPDGX-L00/COB	12 x 20,2 ml R1 + 12 x 18,3 ml R2	1800			

CALIBRATORS

KIT 5 LEV	CPRLK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CRP XL WIDE RANGE

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CPTXL-B00/COB	1 x 24,5 ml R1 + 1 x 15,8 ml R1 + 1 x 7,2 ml R2	189	36 months	serum or plasma	0.607 to 425 (mg/l)
	CPTXL-H00/COB	3 x 24,5 ml R1 + 3 x 15,8 ml R1 + 3 x 7,2 ml R2	566			
	CPTXL-L00/COB	12 x 24,5 ml R1 + 12 x 15,8 ml R1 + 12 x 7,2 ml R2	2264			

CALIBRATORS

KIT 5 LEV	CPREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CYSTATIN C - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CYCOL-B00/COB	1 x 23,5 ml R1 + 1 x 10,5 ml R2	100	6 months	serum or plasma	0.15 to 9.355 (mg/l)
	CYCOL-H00/COB	2 x 23,5 ml R1 + 2 x 10,5 ml R2	200			
	CYCOL-L00/COB	12 x 23,5 ml R1 + 12 x 10,5 ml R2	1200			

CALIBRATORS

KIT 5 LEV	CYREK-000	5x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

H	CYCOX-002	1x2 ml	12 months
M	CYCON-002	1x2 ml	12 months
L	CYCOS-002	1x2 ml	12 months

CATALOGUE 2019

FERRITIN - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample**	Analytical Range
KIT	FNCOL-B00/COB	R1 1 x 21,6 ml R1 + 1 x 11,7 ml R1 + 1 x 7,7 ml R2	250 500	12 months	serum	6.344 to 500 (µg/l)
	FNCOL-H00/COB	R2 2 x 21,6 ml R1 + 2 x 11,7 ml R1 + 2 x 7,7 ml R2				

CALIBRATORS

KIT 5 LEV	FNRGK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	FNCON-002	1x2 ml	18 months
---	-----------	--------	-----------

FIBRINOGEN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	FITUR-B00/COB	R1 1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9 ml R2	250 500	36 months	plasma (from sodium citrate)	0.05 to 10 (g/l)
	FITUR-H00/COB	R2 2 x 24,3 ml R1 + 2 x 15,7 ml R1 + 2 x 9 ml R2				

CALIBRATORS

M	FIREM-001	1x1 ml	12 months
---	-----------	--------	-----------

CONTROLS

M	FICON-001	1x1 ml	12 months
---	-----------	--------	-----------

HAPTOGLOBIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	HATUR-B00/COB	R1 1 x 22,2 ml R1 + 1 x 14,3 ml R1 + 1 x 8,3 ml R2	150 300	36 months	serum	0.002 to 3.86 (g/l)
	HATUR-H00/COB	R2 2 x 22,2 ml R1 + 2 x 14,3 ml R1 + 2 x 8,3 ml R2				

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE 2019

IMMUNOGLOBULINS A

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IATUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2		186	36 months	serum	0.02 to 8.0 (g/l)
	IATUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2		558			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS G

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGTUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 14,9 ml R2		187	36 months	serum	0.125 to 35 (g/l)
	IGTUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 14,9 ml R2		560			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS G (CSF)

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGLCR-B00/COB	1 x 23,5 ml R1 + 1 x 12 ml R1 + 1 x 7,7 ml R2		200	24 months	serum	3,375 to 249,78 (mg/l)
	IGLCR-H00/COB	2 x 23,5 ml R1 + 2 x 12 ml R1 + 2 x 7,7 ml R2		400			

CALIBRATORS

KIT 5 LEV	IGREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

M	IGCON-002	1x2 ml	24 months
L	IGCOS-002	1x2 ml	24 months

CATALOGUE 2019

IMMUNOGLOBULINS M

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample**	Analytical Range	
KIT	IMTUR-B00/COB IMTUR-H00/COB	R1	R2	186 558	36 months	serum	0.038 to 7.0 (g/l)
		1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2	1 x 15,7 ml R1 + 1 x 9,1 ml R2				
		3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2	3 x 9,1 ml R2				

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

LIPOPROTEIN (A)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
KIT	LPTUR2-B00/COB LPTUR2-H00/COB	R1	R2	200 400	12 months	serum or plasma	2.35 to 94.60 (mg/dl)
		1 x 23,5 ml R1 + 1 x 17,2 ml R1 + 1 x 6,7 ml R2	1 x 6,7 ml R2				
		2 x 23,5 ml R1 + 2 x 17,2 ml R1 + 2 x 6,7 ml R2	2 x 6,7 ml R2				

CALIBRATORS

KIT 4 LEV	LPREK-000	4x0,5 ml	10 months
-----------	-----------	----------	-----------

CONTROLS

M	LPCON-002	1x2 ml	12 months
---	-----------	--------	-----------

MICROALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
KIT	MATUR-B00/COB MATUR-H00/COB MATUR-L00/COB	R1	R2	230 460 2760	36 months	urine	1.86 to 400 (mg/l)
		1 x 25,3 ml R1 + 1 x 18,2 ml R1 + 1 x 9,8 ml R2	1 x 9,8 ml R2				
		2 x 25,3 ml R1 + 2 x 18,2 ml R1 + 2 x 9,8 ml R2	2 x 9,8 ml R2				
		12 x 25,3 ml R1 + 12 x 18,2 ml R1 + 12 x 9,8 ml R2	12 x 9,8 ml R2				

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOX-002	1x2 ml	18 months
M	MACON-002	1x2 ml	18 months
L	MACOS-002	1x2 ml	18 months

CATALOGUE 2019

PREALBUMIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	PATUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2		186	36 months	serum and plasma	6.53 to 1000 (mg/l)
	PATUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2		558			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

RETINOL BINDING PROTEIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	RBTUR-B00/COB	1 x 20,7 ml R1 + 1 x 7,7 ml R2		200	24 months	urine	7,81 to 141 (mg/l)
	RBTUR-H00/COB	2 x 20,7 ml R1 + 2 x 7,7 ml R2		400			
	RBTUR-L00/COB	12 x 20,7 ml R1 + 12 x 7,7 ml R2		2400			

CALIBRATORS

KIT 5 LEV	RBREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RBCOX-002	1x2 ml	24 months
M	RBCON-002	1x2 ml	24 months
L	RBCOS-002	1x2 ml	24 months

RHEUMATOID FACTOR (RF) - Colloidal Gold

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	RFCOL-B00/COB	1 x 18,2 ml R1 + 1 x 9,9 ml R1 + 1 x 16,7 ml R2		200	18 months	serum and plasma	2.228 to 120 (kU/l)
	RFCOL-H00/COB	3 x 18,2 ml R1 + 3 x 9,9 ml R1 + 3 x 16,7 ml R2		600			

CALIBRATORS

KIT 5 LEV	RFREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

M	RFCON-002	1x2 ml	24 months
---	-----------	--------	-----------

CATALOGUE 2019

TRANSFERRIN

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT		<i>R1</i>	<i>R2</i>			
	TFTUR-B00/COB	1 x 24,5 ml R1 + 1 x 16,7 ml R1 + 1 x 12,5 ml R2	180	36 months	serum and plasma	0.076 to 8.0 (g/l)
	TFTUR-H00/COB	1 x 24,5 ml R1 + 1 x 16,7 ml R1 + 1 x 12,5 ml R2	500			
TFTUR-L00/COB	1 x 24,5 ml R1 + 1 x 16,7 ml R1 + 1 x 12,5 ml R2	2160				

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months



Your expert in specific proteins

(*) Tests : Amount of real tests. May vary from a analyzer to another one.

() Shelf Life :** DiAgam provides to distributors a minimum of **12 months** (except for Cystatin-c, Ferritin and Lp(a): < 12 months). The time period written in this document shows the maximum shelf life time after manufacturing.

(*) Sample :** DiAgam may provide some applications for other samples, such as LCR, Urin, Veterinary samples etc.... Please contact your distributor or us directly at mail@diagam.com or dial +32 68 55 14 82.

(**) Analytical range :** For general information purposes only. More details can be found in the package insert (instructions for use). Please contact your distributor or us directly at mail@diagam.com or dial +32 68 55 14 82 in order to get the updated data.



Medical device for in vitro diagnosis.

DIAGAM SA, Rue du Parc Industriel 40, 7822 Ghislenghien, BELGIUM



The products fulfill the requirements of European directive 98/79 EC concerning the medical devices of in vitro diagnosis.