

DiAgam

CATALOGUE

ARCHITECT Line





Company profile

MISSION STATEMENT

«To have a perfect control of Development, Manufacturing process & Selling of In Vitro Diagnostic liquid reagents with a strong focus on our Customer vision»

Your R&D Partner

DiAgam manufactures **reagents for immunoturbidimetry**.

Our company specializes in high technology R&D, based on immunology technologies. DiAgam is the creator of liquid laboratory reagents containing **gold particles**.

We are currently taking part in a project to develop **new nanotechnologies** based on the use of **colloidal particles** in the field of medical diagnosis. This activity has also succeeded in completing our products range with Cystatin C, Ferritin & Rheumatoid factor.

Manufacturer of ready-to-use reagents

All our products are developed and manufactured within the company.

Our reagents, controls and calibrators are liquid, stable and ready to use on **all opened biochemistry analyzers even with barcodes** (Abbott - Alinity – Architect , Beckman Coulter – AU series, IDS-iSYS Multi-Discipline Automated System, Menarini – Zenit RA, Mindray BS series, Ortho - VITROS® 5,1 FS Chemistry, VITROS® 4600 Chemistry, VITROS® 5600 Integrated Systems, Roche - Modular®-COBAS C, Siemens - Advia® 1800 - Advia® 2400 - Atellica®, Thermo-Konelab®, etc...)

DiAgam is a member of the National Association of **European Diagnostic Manufacturers Association (EDMA)**. All reagents are labelled and DiAgam is certified **ISO 9001:2015** and **ISO 13485:2016**.

All calibrators and controls are traceable to the **International Standards** (WHO, IFCC, NIBSC, etc.)

Worldwide reliability

DiAgam promotes its products in B to B and B to C businesses through its direct affiliates in Belgium, France and Iberia. We also export in ASEAN, EMAE, Americas.

Thanks to our **manufacturing flexibility**, we can propose all our parameters in three different packagings: **Kits** especially dedicated to particular analyzers, **Vials** filled and **Bulk**, available under **DiAgam** label or in **OEM**.

We respect our philosophy by offering our customer, a **high quality of our products** with a most **competitive price** along with a continuous quest to **satisfy their needs**.




MENU ASSAYS

ALBUMIN
ALPHA 1 - ACID GLYCOPROTEIN
ALPHA 1 - ANTITRYPSIN
ALPHA 1 - MICROGLOBULIN
ALPHA 2 - MACROGLOBULIN
ASO
ANTITHROMBIN III
APOLIPOPROTEIN A1
APOLIPOPROTEIN B
CERULOPLASMIN
BETA 2 MICROGLOBULIN
C3
C4
CRP XL WIDE RANGE

CRP Gold Latex
CYSTATIN C - Colloidal Gold
FERRITIN - Colloidal Gold
FIBRINOGEN
HAPTOGLOBIN
Ig A
Ig E
Ig G
Ig M
LIPOPROTEIN A
MICROALBUMIN
PREALBUMIN
RETINOL BINDING PROTEIN
RHEUMATOID FACTOR - Colloidal Gold
TRANSFERRIN





CATALOGUE

ALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ALTUR-B00/ARC	1 x 40 ml R1 + 1 x 10 ml R2		169	36 months	serum or plasma	0.40 to 121.71 (g/l)
	ALTUR-H00/ARC	1 x 80 ml R1 + 1 x 20 ml R2		372			
	ALTUR-L00/ARC	7 x 80 ml R1 + 7 x 20 ml R2		2604			
CALIBRATORS							
KIT 5 LEV	MPREK-000	5x1 ml			24 months		
CONTROLS							
H	MPCOX-002	1x2 ml			24 months		
M	MPCON-002	1x2 ml			24 months		
L	MPCOS-002	1x2 ml			24 months		
ALBUMIN CSF							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	MATUR-B00/ARC	1 x 22,5 ml R1 + 1 x 6,1 ml R2		100	36 months	CSF	1.86 to 400 (mg/l)
	MATUR-H00/ARC	1 x 74 ml R1 + 1 x 17,5 ml R2		361			
	MATUR-L00/ARC	7 x 75 ml R1 + 7 x 15 ml R2		2126			
CALIBRATORS							
KIT 5 LEV	MAREK-000	5x1 ml			18 months		
CONTROLS							
H	MACOZ-002	1 x 2 ml			18 months		
H	MACOX-002	1x2 ml			18 months		
M	MACON-002	1x2 ml			18 months		
L	MACOS-002	1x2 ml			18 months		
ALPHA-1 GLYCOPROTEIN ACIDE/OROSOMUCOID							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	GPTUR-B00/ARC	1 x 37,5 ml R1 + 1 x 7,5 ml R2		97	36 months	serum	0.143 to 2.42 (g/l)
	GPTUR-H00/ARC	2 x 75 ml R1 + 2 x 15 ml R2		443			
CALIBRATORS							
KIT 5 LEV	MPREK-000	5x1 ml			24 months		
CONTROLS							
H	MPCOX-002	1x2 ml			24 months		
M	MPCON-002	1x2 ml			24 months		
L	MPCOS-002	1x2 ml			24 months		

CATALOGUE

ALPHA-1 ANTITRYPSIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	ATTUR-B00/ARC	R1 1 x 45 ml R1 + 1 x 9 ml R2	122	24 months	serum or plasma	0.007 to 3.5 (g/l)
	ATTUR-H00/ARC	R2 2 x 75 ml R1 + 2 x 15 ml R2	443			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	A1DGX-B00/ARC	R1 1 x 15 ml R1 + 1 x 15 ml R2	150	6 months	serum or plasma	1,29 to 166,6 (mg/l)
	A1DGX-H00/ARC	R2 1 x 55 ml R1 + 1 x 55 ml R2	456			

CALIBRATORS

KIT 5 LEV	A1REK-000	5x1 ml	6 months
-----------	-----------	--------	----------

CONTROLS

M	A1CON-002	1x2 ml	6 months
L	A1COS-002	1x2 ml	6 months

ALPHA-2 MACROGLOBULIN - ALPHA-2 MACROGLOBULIN URINARY

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	MGTUR-H00/ARC	R1 1 x 55 ml R1 + 1 x 11 ml R2	155	36 months	serum or plasma or urine	0.026 to 4.96 (g/l)
	MGTUR-L00/ARC	R2 10 x 55 ml R1 + 10 x 11 ml R2	1550			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months
URINARY LOW	MGCU1-002	1x2 ml	12 months
URINARY HIGH	MGCU2-002	1x2 ml	12 months

ANTI STREPTOLYSIN O (ASO)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	ASLTX-B00/ARC	R1 1 x 70 ml R1 + 1 x 8 ml R2	296	20 months	serum or plasma	19.99 to 864.85 (IU/ml)
	ASLTX-H00/ARC	R2 2 x 87 ml R1 + 2 x 10 ml R2	741			

CALIBRATORS

H	ASREK-000	1x5 ml	12 months
---	-----------	--------	-----------

CONTROLS

M	ASCON-002	1x2 ml	12 months
---	-----------	--------	-----------

CATALOGUE

ANTITHROMBIN III

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT	A3TUR-B00/ARC	R1 1 x 50 ml R1 + 1 x 10 ml R2	169	36 months	plasma	28.15 to 806.30 (mg/l)
	A3TUR-H00/ARC	R2 2 x 70 ml R1 + 2 x 14 ml R2	500			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

APOLIPOPROTEIN A1

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	AATUR-B00/ARC	R1 1 x 38 ml R1 + 1 x 10 ml R2	107	36 months	serum or plasma	0.029 to 400 (mg/dl)
	AATUR-H00/ARC	R2 2 x 75 ml R1 + 2 x 20 ml R2	474			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x0.5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

APOLIPOPROTEIN B

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	ABTUR-B00/ARC	R1 1 x 41 ml R1 + 1 x 8 ml R2	151	36 months	serum or plasma	0.019 to 4.0 (g/l)
	ABTUR-H00/ARC	R2 2 x 67 ml R1 + 2 x 13 ml R2	543			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x0.5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

BETA 2 MICROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	B2DGX-B00/ARC	R1 1 x 27 ml R1 + 1 x 12 ml R2	111	15 months	serum or plasma	0,132 to 95 (mg/l)
	B2DGX-H00/ARC	R2 2 x 45 ml R1 + 2 x 20 ml R2	396			
	B2DGX-L00/ARC	10 x 45 ml R1 + 10 x 20 ml R2	1981			

CALIBRATORS

KIT 5 LEV	B2REK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	B2COX-001	1 x 1 ml	18 months
M	B2CON-003	1x3 ml	18 months
L	B2COS-003	1x3 ml	18 months

CATALOGUE

BETA 2 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	BUDGX-B00/ARC	R1 1 x 23 ml R1 + 1 x 10 ml R2	117	15 months	Urine	0.055 to 9.3 (mg/l)
	BUDGX-H00/ARC	R2 2 x 23 ml R1 + 2 x 10 ml R2	234			

CALIBRATORS

KIT 5 LEV	BUREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	B2CON-003	1x3 ml	18 months
L	B2COS-003	1x3 ml	18 months

CERULOPLASMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	CETUR-B00/ARC	R1 1 x 40 ml R1 + 1 x 8 ml R2	128	36 months	serum	9.48 to 1051.20 (mg/l)
	CETUR-H00/ARC	R2 2 x 60 ml R1 + 2 x 12 ml R2	419			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C3

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	C3TUR-B00/ARC	R1 1 x 41 ml R1 + 1 x 8 ml R2	137	36 months	serum or plasma	0.015 to 4.31 (g/l)
	C3TUR-H00/ARC	R2 2 x 56 ml R1 + 2 x 11 ml R2	405			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

COMPLEMENT C4

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT		<i>R1</i>	<i>R2</i>			
	C4TUR-B00/ARC	1 x 40 ml R1 + 1 x 8 ml R2	128	36 months	serum or plasma	0.002 to 0.825 (g/l)
	C4TUR-H00/ARC	2 x 60 ml R1 + 2 x 12 ml R2	419			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CRP GOLD LATEX

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT		<i>R1</i>	<i>R2</i>			
	CPDGX-B00/ARC	1 x 15 ml R1 + 1 x 15 ml R2	105	16 months	serum or plasma	0.031 to 300 (mg/l)
	CPDGX-H00/ARC	1 x 55 ml R1 + 1 x 55 ml R2	456			
CPDGX-L00/ARC	10 x 55 ml R1 + 10 x 55 ml R2	4561				

CALIBRATORS

KIT 5 LEV	CPRLK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CRP XL WIDE RANGE

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT		<i>R1</i>	<i>R2</i>			
	CPTXL-B00/ARC	1 x 55,5 ml R1 + 1 x 8 ml R2	140	36 months	serum or plasma	0.607 to 425 (mg/l)
	CPTXL-H00/ARC	2 x 85 ml R1 + 2 x 12 ml R2	459			
CPTXL-L00/ARC	7 x 85 ml R1 + 7 x 12 ml R2	1607				

CALIBRATORS

KIT 5 LEV	CPREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CATALOGUE

CYSTATIN C - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT		<i>R1</i> <i>R2</i>				
	CYCOL-B00/ARC	1 x 27 ml R1 + 1 x 11 ml R2	107	6 months	serum or plasma	0.15 to 9.355 (mg/l)
	CYCOL-H00/ARC	2 x 48 ml R1 + 2 x 20 ml R2	421			
CYCOL-L00/ARC	10 x 48 ml R1 + 10 x 20 ml R2	2103				

CALIBRATORS

KIT 5 LEV	CYREK-000	5x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

H	CYCOX-002	1x2 ml	12 months
M	CYCON-002	1x2 ml	12 months
L	CYCOS-002	1x2 ml	12 months

FERRITIN - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample**	Analytical Range
KIT		<i>R1</i> <i>R2</i>				
	FNCOL-B00/ARC	1 x 40 ml R1 + 1 x 8 ml R2	128	12 months	serum	6.344 to 500 (µg/l)
FNCOL-H00/ARC	2 x 55 ml R1 + 2 x 11 ml R2	378				

CALIBRATORS

KIT 5 LEV	FNRGK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	FNCON-002	1x2 ml	18 months
---	-----------	--------	-----------

FIBRINOGEN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT		<i>R1</i> <i>R2</i>				
	FITUR-B00/ARC	1 x 37,5 ml R1 + 1 x 7,5 ml R2	97	36 months	plasma (from sodium citrate)	0.05 to 10 (g/l)
FITUR-H00/ARC	2 x 50 ml R1 + 1 x 15 ml R2	222				

CALIBRATORS

M	FIREM-001	1x1 ml	12 months
---	-----------	--------	-----------

CONTROLS

M	FICON-001	1x1 ml	12 months
---	-----------	--------	-----------

CATALOGUE

HAPTOGLOBIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	HATUR-B00/ARC	1 x 37,5 ml R1 + 1 x 7,5 ml R2	97	36 months	serum	0.002 to 3.86 (g/l)
	HATUR-H00/ARC	2 x 75 ml R1 + 2 x 15 ml R2	443			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS A

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample**	Analytical Range
		<i>R1</i>	<i>R2</i>			
KIT	IATUR-B00/ARC	1 x 38 ml R1 + 1 x 8 ml R2	144	36 months	serum	0.02 to 8.0 (g/l)
	IATUR-H00/ARC	2 x 54 ml R1 + 2 x 11 ml R2	425			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS E

Contact us

IMMUNOGLOBULINS G

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample**	Analytical Range
		<i>R1</i>	<i>R2</i>			
KIT	IGTUR-B00/ARC	1 x 25 ml R1 + 1 x 9,5 ml R2	104	36 months	serum	0.125 to 35 (g/l)
	IGTUR-H00/ARC	2 x 50 ml R1 + 2 x 19 ml R2	461			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

IMMUNOGLOBULINS G (CSF)

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT	IGLCR-B00/ARC	R1 1 x 18,5 ml R1 + 1 x 5 ml R2	100	36 months	serum	3,375 to 249,78 (mg/l)
	IGLCR-H00/ARC	R2 1 x 74 ml R1 + 1 x 17 ml R2	464			

CALIBRATORS

KIT 5 LEV	IGREK-000	5x1 ml	24 months
------------------	-----------	--------	-----------

CONTROLS

H	IGCOX-002	1x2 ml	24 months
M	IGCON-002	1x2 ml	24 months
L	IGCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS M

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT	IMTUR-B00/ARC	R1 1 x 38 ml R1 + 1 x 8 ml R2	144	36 months	serum	0.038 to 7.0 (g/l)
	IMTUR-H00/ARC	R2 2 x 49,5 ml R1 + 2 x 10,5 ml R2	402			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
------------------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

LIPOPROTEIN (A)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	LPTUR2-B00/ARC	R1 1 x 64 ml R1 + 1 x 8 ml R2	191	12 months	serum or plasma	2.35 to 94.60 (mg/dl)
	LPTUR2-H00/ARC	R2 2 x 80 ml R1 + 2 x 10 ml R2	503			

CALIBRATORS

KIT 4 LEV	LPREK-000	4x0,5 ml	10 months
------------------	-----------	----------	-----------

CONTROLS

M	LPCON-002	1x2 ml	12 months
----------	-----------	--------	-----------

CATALOGUE

MICROALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	MATUR-B00/ARC	1 x 22,5 ml R1 + 1 x 6,1 ml R2		100	36 months	urine	1.86 to 400 (mg/l)
	MATUR-H00/ARC	1 x 74 ml R1 + 1 x 17,5 ml R2		361			
	MATUR-L00/ARC	7 x 75 ml R1 + 7 x 15 ml R2		2126			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOX-002	1x2 ml	18 months
M	MACON-002	1x2 ml	18 months
L	MACOS-002	1x2 ml	18 months

PREALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	PATUR-H00/ARC	1 x 55 ml R1 + 1 x 11 ml R2		155	36 months	serum and plasma	6.53 to 1000 (mg/l)

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

RETINOL BINDING PROTEIN (RBP)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	RBTUR-B00/ARC	1 x 31 ml R1 + 1 x 10 ml R2		138	24 months	urine	7,81 to 141 (mg/l)
	RBTUR-H00/ARC	2 x 31 ml R1 + 2 x 10 ml R2		277			
	RBTUR-L00/ARC	10 x 31 ml R1 + 10 x 10 ml R2		1383			

CALIBRATORS

KIT 5 LEV	RBREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RBCOX-002	1x2 ml	24 months
M	RBCON-002	1x2 ml	24 months
L	RBCOS-002	1x2 ml	24 months

CATALOGUE

RHEUMATOID FACTOR (RF) - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	RFCOL-B00/ARC	R1 1 x 30 ml R1 + 1 x 20 ml R2	122	18 months	serum and plasma	2.228 to 120 (kUI/l)
	RFCOL-H00/ARC	R2 2 x 60 ml R1 + 2 x 40 ml R2	512			

CALIBRATORS

KIT 5 LEV	RFREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

M	RFCON-002	1x2 ml	24 months
---	-----------	--------	-----------

TRANSFERRIN

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT	TFTUR-B00/ARC	R1 1 x 37,5 ml R1 + 1 x 10,5 ml R2	108	36 months	serum and plasma	0.076 to 8.0 (g/l)
	TFTUR-H00/ARC	R2 2 x 72 ml R1 + 2 x 20 ml R2	449			
	TFTUR-L00/ARC	7 x 72 ml R1 + 7 x 20 ml R2	1570			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months



Your expert in specific proteins

(*) Tests : Amount of real tests. May vary from a analyzer to another one.

() Shelf Life :** DiAgam provides to distributors a minimum of **12 months** (except for Cystatin-c, Ferritin and Lp(a): < 12 months). The time period written in this document shows the maximum shelf life time after manufacturing.

(*) Sample :** DiAgam may provide some applications for other samples, such as LCR, Urin, Veterinary samples etc.... Please contact your distributor or us directly at mail@diagam.com or dial +32 68 55 14 82.

(**)**
Analytical range : For general information purposes only. More details can be found in the package insert (instructions for use). Please contact your distributor or us directly at mail@diagam.com or dial +32 68 55 14 82 in order to get the updated data.

IVD Medical device for in vitro diagnosis.

 DIAGAM SA, Rue du Parc Industriel 40, 7822 Ghislenghien, BELGIUM

CE The products fulfill the requirements of European directive 98/79 EC concerning the medical devices of in vitro diagnosis.

© DiAgam. 2020. All rights reserved. All product names, registered trademarks, company names in this document remain the property of their respective owners. For general information purposes only. More details can be found in the package insert (instructions for use) and labelling. Not for the US market except for research use only.