



## CATALOGUE 2019

### *ARCHITECT Line*





## Company profile

### MISSION STATEMENT

**«To have a perfect control of Development, Manufacturing process & Selling of In Vitro Diagnostic liquid reagents with a strong focus on our Customer vision»**

#### Your R&D Partner

DiAgam manufactures **reagents for immunoturbidimetry**.

Our company specializes in high technology R&D, based on immunology technologies. DiAgam is the creator of liquid laboratory reagents containing **gold particles**.

We are currently taking part in a project to develop **new nanotechnologies** based on the use of **colloidal particles** in the field of medical diagnosis. This activity has also succeeded in completing our products range with Cystatin C, Ferritin & Rheumatoid factor.

#### Manufacturer of ready-to-use reagents

All our products are developed and manufactured within the company.

Our reagents, controls and calibrators are liquid, stable and ready to use on **all opened biochemistry ana-lyzers even with barcodes** (Abbott - Alinity – Architect , Beckman Coulter – AU series, IDS-iSYS Multi-Discipline Automated System, Menarini – Zenit RA, Mindray BS series, Ortho - VITROS® 5,1 FS Chemistry, VITROS® 4600 Chemistry, VITROS® 5600 Integrated Systems, Roche - Modular®- COBAS C, Siemens - Advia® 1800 - Advia® 2400 - Atellica®, Thermo-Konelab®, etc...)

DiAgam is a member of the National Association of **European Diagnostic Manufacturers Association (EDMA)**. All reagents are labelled and DiAgam is certified **ISO 9001:2015** and **ISO 13485:2016**.

All calibrators and controls are traceable to the **International Standards** (WHO, IFCC, NIBSC, etc.)

#### Worldwide reliability

DiAgam is present all around the world. We sell directly in Belgium, France, Spain & Portugal. We also export to French Speaking African countries, Argentina, Brazil, Czech Republic, Greece, Hungary, India, Mexico, Morocco, Poland, Sri Lanka, Ukraine, Vietnam, etc.

Thanks to our **manufac-turing flexibility**, we can propose all our parameters in three different packagings: **Kits** especially dedicated to particular analyzers, **Vials** filled and **Bulk**, available under **DiAgam** label or in **OEM**.

We respect our philosophy by offering our cus-tomer, a **high quality of our products** with a most **competitive price** along with a continuous quest to **satisfy their needs**.

# CATALOGUE 2019

## ALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	ALTUR-B00/ARC	1 x 40 ml R1 + 1 x 10 ml R2	250	36 months	serum or plasma	0.40 to 121.71 (g/l)
	ALTUR-H00/ARC	1 x 80 ml R1 + 1 x 20 ml R2	500			
	ALTUR-L00/ARC	7 x 80 ml R1 + 7 x 20 ml R2	3500			

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

### CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

## ALPHA-1 GLYCOPROTEIN ACIDE/OROSOMUCOID

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	GPTUR-B00/ARC	1 x 37,5 ml R1 + 1 x 7,5 ml R2	150	36 months	serum	0.143 to 2.42 (g/l)
	GPTUR-H00/ARC	2 x 75 ml R1 + 2 x 15 ml R2	600			

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

### CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

## ALPHA-1 ANTITRYPSIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	ATTUR-B00/ARC	1 x 45 ml R1 + 1 x 9 ml R2	180	24 months	serum or plasma	0.007 to 3.5 (g/l)
	ATTUR-H00/ARC	2 x 75 ml R1 + 2 x 15 ml R2	800			

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

### CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

## ALPHA-1 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT						Contact us

# CATALOGUE 2019

## ALPHA-2 MACROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	MGTUR-H00/ARC	1 x 55 ml R1 + 1 x 11 ml R2		220	36 months	serum or plasma	0.026 to 4.96 (g/l)
	MGTUR-L00/ARC	10 x 55 ml R1 + 10 x 11 ml R2		2200			

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

### CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months
URINARY LOW	MGCU1-002	1x2 ml	12 months
URINARY HIGH	MGCU2-002	1x2 ml	12 months

## ANTI STREPTOLYSIN O (ASO)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	ASLTX-B00/ARC	1 x 70 ml R1 + 1 x 8 ml R2		296	20 months	serum or plasma	19.99 to 864.85 (IU/ml)
	ASLTX-H00/ARC	2 x 87 ml R1 + 2 x 10 ml R2		741			

### CALIBRATORS

H	ASREH-001(-005)	1x1 or (1x5) ml	12 months
---	-----------------	-----------------	-----------

### CONTROLS

M	ASCON-002	1x2 ml	12 months
---	-----------	--------	-----------

## ANTITHROMBIN III

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range	
		R1	R2				
KIT	A3TUR-B00/ARC	1 x 50 ml R1 + 1 x 10 ml R2		250	36 months	plasma	28.15 to 806.30 (mg/l)
	A3TUR-H00/ARC	2 x 70 ml R1 + 2 x 15 ml R2		700			

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

### CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

# CATALOGUE 2019

## APOLIPOPROTEIN A1

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	AATUR-B00/ARC	1 x 38 ml R1 + 1 x 10 ml R2		152	36 months	serum or plasma	0.029 to 400 (mg/dl)
	AATUR-H00/ARC	2 x 75 ml R1 + 2 x 20 ml R2		600			

### CALIBRATORS

KIT 5 LEV	ABREK-000	5x0.5 ml	18 months
-----------	-----------	----------	-----------

### CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

## APOLIPOPROTEIN B

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	ABTUR-B00/ARC	1 x 41 ml R1 + 1 x 8 ml R2		242	36 months	serum or plasma	0.019 to 4.0 (g/l)
	ABTUR-H00/ARC	2 x 67 ml R1 + 2 x 13 ml R2		788			

### CALIBRATORS

KIT 5 LEV	ABREK-000	5x0.5 ml	18 months
-----------	-----------	----------	-----------

### CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

## BETA 2 MICROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	B2DGX-B00/ARC	1 x 27 ml R1 + 1 x 12 ml R2		150	15 months	serum or plasma	0,132 to 95 (mg/l)
	B2DGX-H00/ARC	2 x 45 ml R1 + 2 x 20 ml R2		500			
	B2DGX-L00/ARC	10 x 45 ml R1 + 10 x 20 ml R2		2500			

### CALIBRATORS

KIT 5 LEV	B2REK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

### CONTROLS

H	B2COX-001	1 x 1 ml	18 months
M	B2CON-003	1x3 ml	18 months
L	B2COS-003	1x3 ml	18 months

## BETA 2 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
<b>KIT</b>	BUDGX-B00/ARC BUDGX-H00/ARC BUDGX-L00/ARC					
		<b>Contact us</b>		15 months	Urine	0.055 to 9.3 (mg/l)

### CALIBRATORS

KIT 5 LEV	BUREK-000	5x1 ml		18 months		
-----------	-----------	--------	--	-----------	--	--

### CONTROLS

M	B2CON-003	1x3 ml		18 months		
L	B2COS-003	1x3 ml		18 months		

## CERULOPLASMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
<b>KIT</b>	CETUR-B00/ARC CETUR-H00/ARC	1 x 40 ml R1 + 1 x 8 ml R2 2 x 60 ml R1 + 2 x 12 ml R2	200 600	36 months	serum	9.48 to 1051.20 (mg/l)

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

### CONTROLS

H	MPCOX-002	1x2 ml		24 months		
M	MPCON-002	1x2 ml		24 months		
L	MPCOS-002	1x2 ml		24 months		

## COMPLEMENT C3

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
<b>KIT</b>	C3TUR-B00/ARC C3TUR-H00/ARC	1 x 41 ml R1 + 1 x 8 ml R2 2 x 56 ml R1 + 2 x 11 ml R2	216 595	36 months	serum or plasma	0.015 to 4.31 (g/l)

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

### CONTROLS

H	MPCOX-002	1x2 ml		24 months		
M	MPCON-002	1x2 ml		24 months		
L	MPCOS-002	1x2 ml		24 months		

## COMPLEMENT C4

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT	C4TUR-B00/ARC	R1 1 x 40 ml R1 + 1 x 8 ml R2	200	36 months	serum or plasma	0.002 to 0.825 (g/l)
	C4TUR-H00/ARC	R2 2 x 60 ml R1 + 2 x 12 ml R2				

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

### CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

## CRP GOLD LATEX

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
KIT	CPDGX-B00/ARC	R1 1 x 15 ml R1 + 1 x 15 ml R2	150	16 months	serum or plasma	0.031 to 300 (mg/l)	
	CPDGX-H00/ARC	R2 1 x 55 ml R1 + 1 x 55 ml R2					550
	CPDGX-L00/ARC	10 x 55 ml R1 + 10 x 55 ml R2					5500

### CALIBRATORS

KIT 5 LEV	CPRLK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

### CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

## CRP XL WIDE RANGE

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
KIT	CPTXL-B00/ARC	R1 1 x 55,5 ml R1 + 1 x 8 ml R2	222	36 months	serum or plasma	0.607 to 425 (mg/l)	
	CPTXL-H00/ARC	R2 2 x 85 ml R1 + 2 x 12 ml R2					667
	CPTXL-L00/ARC	7 x 85 ml R1 + 7 x 12 ml R2					2333

### CALIBRATORS

KIT 5 LEV	CPREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

### CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

# CATALOGUE 2019

## CYSTATIN C - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
KIT	CYCOL-B00/ARC	1 x 27 ml R1 + 1 x 11 ml R2	147	6 months	serum or plasma	0.15 to 9.355 (mg/l)
	CYCOL-H00/ARC	2 x 48 ml R1 + 2 x 20 ml R2	533			
	CYCOL-L00/ARC	10 x 48 ml R1 + 10 x 20 ml R2	2667			

### CALIBRATORS

KIT 5 LEV	CYREK-000	5x1 ml	12 months
-----------	-----------	--------	-----------

### CONTROLS

H	CYCOX-002	1x2 ml	12 months
M	CYCON-002	1x2 ml	12 months
L	CYCOS-002	1x2 ml	12 months

## FERRITIN - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		R1 R2				
KIT	FNCOL-B00/ARC	1 x 40 ml R1 + 1 x 8 ml R2	200	12 months	serum	6.344 to 500 (µg/l)
	FNCOL-H00/ARC	2 x 55 ml R1 + 2 x 11 ml R2	550			

### CALIBRATORS

KIT 5 LEV	FNRGK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

### CONTROLS

M	FNCON-002	1x2 ml	18 months
---	-----------	--------	-----------

## FIBRINOGEN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
KIT	FITUR-B00/ARC	1 x 37,5 ml R1 + 1 x 7,5 ml R2	150	36 months	plasma (from sodium citrate)	0.05 to 10 (g/l)
	FITUR-H00/ARC	2 x 50 ml R1 + 1 x 15 ml R2	300			

### CALIBRATORS

M	FIREM-001	1x1 ml	12 months
---	-----------	--------	-----------

### CONTROLS

M	FICON-001	1x1 ml	12 months
---	-----------	--------	-----------



# CATALOGUE 2019

## HAPTOGLOBIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i> <i>R2</i>				
<b>KIT</b>	HATUR-B00/ARC HATUR-H00/ARC	1 x 37,5 ml R1 + 1 x 7,5 ml R2 2 x 75 ml R1 + 2 x 15 ml R2	150 600	36 months	serum	0.002 to 3.86 (g/l)

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

### CONTROLS

H	MPCOX-002	1x2 ml		24 months		
M	MPCON-002	1x2 ml		24 months		
L	MPCOS-002	1x2 ml		24 months		

## IMMUNOGLOBULINS A

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		<i>R1</i> <i>R2</i>				
<b>KIT</b>	IATUR-B00/ARC IATUR-H00/ARC	1 x 38 ml R1 + 1 x 8 ml R2 2 x 54 ml R1 + 2 x 11 ml R2	229 629	36 months	serum	0.02 to 8.0 (g/l)

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

### CONTROLS

H	MPCOX-002	1x2 ml		24 months		
M	MPCON-002	1x2 ml		24 months		
L	MPCOS-002	1x2 ml		24 months		

## IMMUNOGLOBULINS G

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		<i>R1</i> <i>R2</i>				
<b>KIT</b>	IGTUR-B00/ARC IGTUR-H00/ARC	1 x 25 ml R1 + 1 x 9,5 ml R2 2 x 50 ml R1 + 2 x 19 ml R2	148 594	36 months	serum	0.125 to 35 (g/l)

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

### CONTROLS

H	MPCOX-002	1x2 ml		24 months		
M	MPCON-002	1x2 ml		24 months		
L	MPCOS-002	1x2 ml		24 months		

## IMMUNOGLOBULINS G (CSF)

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		<i>R1</i> <i>R2</i>				
<b>KIT</b>						<b>Contact us</b>

# CATALOGUE 2019

## IMMUNOGLOBULINS M

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample**	Analytical Range
KIT	IMTUR-B00/ARC IMTUR-H00/ARC	R1 1 x 38 ml R1 + 1 x 8 ml R2	229 600	36 months	serum	0.038 to 7.0 (g/l)
		R2 2 x 49,5 ml R1 + 2 x 10,5 ml R2				

### CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

### CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

## LIPOPROTEIN (A)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	LPTUR2-B00/ARC LPTUR2-H00/ARC	R1 1 x 64 ml R1 + 1 x 8 ml R2	320 800	12 months	serum or plasma	2.35 to 94.60 (mg/dl)
		R2 2 x 80 ml R1 + 2 x 10 ml R2				

### CALIBRATORS

KIT 4 LEV	LPREK-000	4x0,5 ml	10 months
-----------	-----------	----------	-----------

### CONTROLS

M	LPCON-002	1x2 ml	12 months
---	-----------	--------	-----------

## MICROALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	MATUR-B00/ARC MATUR-H00/ARC MATUR-L00/ARC	R1 1 x 40 ml R1 + 1 x 8 ml R2	229 600 3000	36 months	urine	1.86 to 400 (mg/l)
		R2 2 x 52,5 ml R1 + 2 x 10,5 ml R2				
		7 x 75 ml R1 + 7 x 15 ml R2				

### CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

### CONTROLS

H	MACOX-002	1x2 ml	18 months
M	MACON-002	1x2 ml	18 months
L	MACOS-002	1x2 ml	18 months

# CATALOGUE 2019

## PREALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
<b>KIT</b>	PATUR-H00/ARC	1 x 55 ml R1 + 1 x 11 ml R2	220	36 months	serum and plasma	6.53 to 1000 (mg/l)

## CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

## CONTROLS

H	MPCOX-002	1x2 ml		24 months		
M	MPCON-002	1x2 ml		24 months		
L	MPCOS-002	1x2 ml		24 months		

## RETINOL BINDING PROTEIN (RBP)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
<b>KIT</b>			Contact us			

## RHEUMATOID FACTOR (RF) - Colloidal Gold

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1 R2				
<b>KIT</b>	RFCOL-B00/ARC	1 x 30 ml R1 + 1 x 20 ml R2	150	18 months	serum and plasma	2.228 to 120 (kU/l)
	RFCOL-H00/ARC	2 x 60 ml R1 + 2 x 40 ml R2	600			

## CALIBRATORS

KIT 5 LEV	RFREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

## CONTROLS

M	RFCON-002	1x2 ml		24 months		
---	-----------	--------	--	-----------	--	--

## TRANSFERRIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample**	Analytical Range
		R1 R2				
<b>KIT</b>	TFTUR-B00/ARC	1 x 37,5 ml R1 + 1 x 10,5 ml R2	150	36 months	serum and plasma	0.076 to 8.0 (g/l)
	TFTUR-H00/ARC	2 x 72 ml R1 + 2 x 20 ml R2	571			
	TFTUR-L00/ARC	7 x 72 ml R1 + 7 x 20 ml R2	2000			

## CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml		24 months		
-----------	-----------	--------	--	-----------	--	--

## CONTROLS

H	MPCOX-002	1x2 ml		24 months		
M	MPCON-002	1x2 ml		24 months		
L	MPCOS-002	1x2 ml		24 months		



*Your expert in specific proteins*

---

**(\*) Tests :** Amount of theoretical tests. May vary from a analyzer to another one.

**(\*\*) Shelf Life :** DiAgam provides to distributors a minimum of **12 months** (except for Cystatin-c, Ferritin and Lp(a): < 12 months). The time period written in this document shows the maximum shelf life time after manufacturing.

**(\*\*\*) Sample :** DiAgam may provide some applications for other samples, such as LCR, Urin, Veterinary samples etc.... Please contact your distributor or us directly at [mail@diagam.com](mailto:mail@diagam.com) or dial **+32 68 55 14 82**.

**(\*\*\*\*)**  
**Analytical range :** For general information purposes only. More details can be found in the package insert (instructions for use). Please contact your distributor or us directly at [mail@diagam.com](mailto:mail@diagam.com) or dial **+32 68 55 14 82** in order to get the updated data.

**IVD** Medical device for in vitro diagnosis.



DIAGAM SA, Rue du Parc Industriel 40, 7822 Ghislenghien, BELGIUM



The products fulfill the requirements of European directive 98/79 EC concerning the medical devices of in vitro diagnosis.